

## Declaration of Conformity

We,

AVA Monitoring AB  
Vädursgatan 6  
SE-412 50 Göteborg  
SWEDEN

declare under our sole responsibility that the measurement instrument

**AvaTrace M90, HDU Version R1A**

to which this declaration relates is in conformity with all the essential requirements of the Radio Equipment Directive 2014/53/EU (RED). The conformity with the essential requirements set out in article 3 of the Radio Equipment Directive has been verified using the following standards:

<b>Safety</b> (RED article 3.1a)	EN 61010-1:2010+A1:2019 EN IEC 62311:2020
<b>EMC</b> (RED article 3.1b)	EN IEC 61326-1:2021 EN 301 489-1 V2.2.3 EN 301 489-52 V1.2.1
<b>Spectrum</b> (RED article 3.2)	EN 301 511 V12.5.1 EN 301 908-1 V15.1.1 EN 301 908-2 V13.1.1 EN 301 908-13 V13.2.1

The conformity assessment procedure set out in article 17 and Annex II of the Radio Equipment Directive has been followed in the assessment of the essential requirements.

The technical documentation defined in Annex II of the RED Directive is held at

AVA Monitoring AB  
Vädursgatan 6  
SE-412 50 Göteborg  
SWEDEN

Göteborg, 9 October 2024



Markus Leufkens  
CEO, AVA Monitoring AB